

Fox Valley Metrology

3114 Medalist Drive
 Oshkosh, WI 54902
 (920) 426-5894 • Fax (920) 426-8120
 http://www.FoxValleyMetrology.com

CERTIFICATE OF CALIBRATION



Certificate No. ACT-1272

<p>CERTIFICATION NUMBER CL096-32665-430</p> <p>FOR Intertek 8431 Murphy Drive Middleton, WI 53562</p> <p>PURCHASE ORDER #</p> <p>TEST INSTRUMENT Scales</p> <p>MAKE GSE Manufacturing</p> <p>MODEL 450-100</p> <p>SERIAL NUMBER 101722</p> <p>IDENTIFICATION 008 </p> <p>CUSTOMER LOCATION</p> <p>CONDITION RECEIVED In Tolerance</p> <p>CONDITION RETURNED In Tolerance</p> <p>CALIBRATED BY Danny Scherr</p> <p>CALIBRATION LOCATION On Site</p> <p>ENVIRONMENT 70.0°F, 21.1°C, 36.0%RH</p> <p>CALIBRATION DATE 04/06/2021</p> <p>RECALIBRATION DUE 10/06/2021</p>	<p>PROCEDURES FOLLOWED FVE-020 rev. 2</p> <p>STANDARDS USED</p> <table border="1"> <thead> <tr> <th>INSTRUMENT</th> <th>SERIAL NUMBER</th> <th>TRACE NUMBER</th> <th>NEXT CAL</th> </tr> </thead> <tbody> <tr> <td>FVS-019D</td> <td>26623</td> <td>CK240-48617-466</td> <td>08/31/2022</td> </tr> <tr> <td>FVS-019G-ORANGE</td> <td>N/A</td> <td>CK205-68362-614</td> <td>07/31/2021</td> </tr> <tr> <td>FVS-469</td> <td>N/A</td> <td>CL084-50197-397</td> <td>03/31/2022</td> </tr> </tbody> </table> <p>Total expanded measurement uncertainties expressed are based on a confidence level of 95%; coverage factor of (k=2). The statement of compliance in this certificate was issued without taking the uncertainty of measurement into consideration. The customer shall assess the results and uncertainty when determining if the results meet their needs. (This is considered "shared responsibility.") Uncertainties expressed in nominal units.</p> <p>The calibrations within the certificate/report are traceable through NIST or another National Metrology Institute to the International System of Units (SI). Calibration was completed in accordance with ISO/IEC 17025:2017, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. Other standards listed upon request.</p>	INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL	FVS-019D	26623	CK240-48617-466	08/31/2022	FVS-019G-ORANGE	N/A	CK205-68362-614	07/31/2021	FVS-469	N/A	CL084-50197-397	03/31/2022
INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL														
FVS-019D	26623	CK240-48617-466	08/31/2022														
FVS-019G-ORANGE	N/A	CK205-68362-614	07/31/2021														
FVS-469	N/A	CL084-50197-397	03/31/2022														

This certificate shall not be altered in any form or reproduced, except in full, without prior written approval from originating lab. These results relate only to the item(s) calibrated.
 Form Revision 7: 07/14/2020

CALIBRATION RESULTS

* DENOTES "OUT OF TOLERANCE"

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY	
Weight	(1b)	(1b)	(1b)	(1b)	(1b)	(1b)	
	1.00	0.89	1.11	1.00	1.00	0.007	
	5.00	4.89	5.11	5.00	5.00	0.006	
	10.00	9.89	10.11	10.00	10.00	0.007	
	20.00	19.89	20.11	20.00	20.00	0.007	
	40.00	39.89	40.11	40.00	40.00	0.081	
	80.00	79.89	80.11	79.99	79.99	0.080	
	100.00	99.89	100.11	100.00	100.00	0.081	
Shift Test	(1b)	(1b)	(1b)	(1b)	(1b)	(1b)	
	Front	10.00	9.89	10.11	10.00	10.00	0.006
	Left	10.00	9.89	10.11	10.00	10.00	0.007
	Back	10.00	9.89	10.11	10.00	10.00	0.006
	Right	10.00	9.89	10.11	10.00	10.00	0.006