

**CERTIFICATE OF
 CALIBRATION**



Certificate No. ACT-1272

<p>CERTIFICATION NUMBER CL193-46466-679</p> <p>FOR Intertek 8431 Murphy Drive Middleton, WI 53562</p> <p>PURCHASE ORDER #</p> <p>TEST INSTRUMENT Stopwatch</p> <p>MAKE Control Company MODEL 1030 SERIAL NUMBER IDENTIFICATION 1378</p> <p>CUSTOMER LOCATION</p> <p>CONDITION RECEIVED In Tolerance CONDITION RETURNED In Tolerance CALIBRATED BY Danny Scherr</p> <p>CALIBRATION LOCATION On Site ENVIRONMENT 70.0°F, 21.1°C, 47.0%RH CALIBRATION DATE 07/12/2021 RECALIBRATION DUE 07/12/2022</p>	<p>PROCEDURES FOLLOWED FVE-033 rev. 1</p> <p>STANDARDS USED</p> <table border="1"> <thead> <tr> <th>INSTRUMENT</th> <th>SERIAL NUMBER</th> <th>TRACE NUMBER</th> <th>NEXT CAL</th> </tr> </thead> <tbody> <tr> <td>FVS-393D</td> <td>N/A</td> <td>CL004-49737-397</td> <td>01/31/2022</td> </tr> <tr> <td>FVS-469</td> <td>N/A</td> <td>CL084-50197-397</td> <td>03/31/2022</td> </tr> </tbody> </table> <p>Total expanded measurement uncertainties expressed are based on a confidence level of 95%; coverage factor of (k=2). The statement of compliance in this certificate was issued without taking the uncertainty of measurement into consideration. The customer shall assess the results and uncertainty when determining if the results meet their needs. (This is considered "shared responsibility.") Uncertainties expressed in nominal units.</p> <p>The calibrations within the certificate/report are traceable through NIST or another National Metrology Institute to the International System of Units (SI). Calibration was completed in accordance with ISO/IEC 17025:2017, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. Other standards listed upon request.</p>	INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL	FVS-393D	N/A	CL004-49737-397	01/31/2022	FVS-469	N/A	CL084-50197-397	03/31/2022
INSTRUMENT	SERIAL NUMBER	TRACE NUMBER	NEXT CAL										
FVS-393D	N/A	CL004-49737-397	01/31/2022										
FVS-469	N/A	CL084-50197-397	03/31/2022										

This certificate shall not be altered in any form or reproduced, except in full, without prior written approval from originating lab. These results relate only to the item(s) calibrated.
 Form Revision 7: 07/14/2020

CALIBRATION RESULTS

* DENOTES "OUT OF TOLERANCE"

FEATURE	NOMINAL	LOWER LIMIT	UPPER LIMIT	AS FOUND	AS LEFT	UNCERTAINTY
	(sec)	(sec)	(sec)	(sec)	(sec)	(sec)
1 minute	60.00	59.64	60.36	60.03	60.03	0.006
	60.00	59.64	60.36	59.92	59.92	0.006
5 minute	300.00	299.64	300.36	300.06	300.06	0.007
	300.00	299.64	300.36	300.05	300.05	0.006