

**CERTIFICATE OF
 CALIBRATION**



Certificate No. ACT-1272

| <p>CERTIFICATION NUMBER CL095-41794-502</p> <p>FOR Intertek 8431 Murphy Drive Middleton, WI 53562</p> <p>PURCHASE ORDER #</p> <p>TEST INSTRUMENT Timer</p> <p>MAKE Cole-Parmer MODEL 94440-10 SERIAL NUMBER 101587800 IDENTIFICATION 1212</p> <p>CUSTOMER LOCATION Hearth</p> <p>CONDITION RECEIVED In Tolerance CONDITION RETURNED In Tolerance CALIBRATED BY Christopher Moore</p> <p>CALIBRATION LOCATION On Site ENVIRONMENT 72.0°F, 22.2°C, 21.0%RH CALIBRATION DATE 04/05/2021 RECALIBRATION DUE 04/05/2022</p> | <p>PROCEDURES FOLLOWED FVE-033 rev. 1</p> <p>STANDARDS USED</p> <table border="1"> <thead> <tr> <th>INSTRUMENT</th> <th>SERIAL NUMBER</th> <th>TRACE NUMBER</th> <th>NEXT CAL</th> </tr> </thead> <tbody> <tr> <td>FVS-553</td> <td>N/A</td> <td>CL022-19908-397</td> <td>01/31/2022</td> </tr> <tr> <td>FVS-811B</td> <td>N/A</td> <td>CK281-24359-397</td> <td>10/31/2021</td> </tr> </tbody> </table> <p>Total expanded measurement uncertainties expressed are based on a confidence level of 95%; coverage factor of (k=2). The statement of compliance in this certificate was issued without taking the uncertainty of measurement into consideration. The customer shall assess the results and uncertainty when determining if the results meet their needs. (This is considered "shared responsibility.") Uncertainties expressed in nominal units.</p> <p>The calibrations within the certificate/report are traceable through NIST or another National Metrology Institute to the International System of Units (SI). Calibration was completed in accordance with ISO/IEC 17025:2017, ANSI/NCSL Z540-1-1994 and ANSI/NCSL Z540.3-2006. Other standards listed upon request.</p> | INSTRUMENT | SERIAL NUMBER | TRACE NUMBER | NEXT CAL | FVS-553 | N/A | CL022-19908-397 | 01/31/2022 | FVS-811B | N/A | CK281-24359-397 | 10/31/2021 |
|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------|---------------|--------------|----------|---------|-----|-----------------|------------|----------|-----|-----------------|------------|
| INSTRUMENT | SERIAL NUMBER | TRACE NUMBER | NEXT CAL | | | | | | | | | | |
| FVS-553 | N/A | CL022-19908-397 | 01/31/2022 | | | | | | | | | | |
| FVS-811B | N/A | CK281-24359-397 | 10/31/2021 | | | | | | | | | | |

CALIBRATION RESULTS

* DENOTES "OUT OF TOLERANCE"

| FEATURE | NOMINAL | LOWER LIMIT | UPPER LIMIT | AS FOUND | AS LEFT | UNCERTAINTY |
|---------|---------|-------------|-------------|----------|---------|-------------|
| Timer | (sec) | (sec) | (sec) | (sec) | (sec) | (sec) |
| | 60.0 | 59.5 | 60.5 | 60.0 | 60.0 | 0.06 |
| | 60.0 | 59.5 | 60.5 | 59.9 | 59.9 | 0.07 |
| | (sec) | (sec) | (sec) | (sec) | (sec) | (sec) |
| | 300.0 | 299.5 | 300.5 | 300.0 | 300.0 | 0.06 |
| | 300.0 | 299.5 | 300.5 | 300.0 | 300.0 | 0.06 |
| | (sec) | (sec) | (sec) | (sec) | (sec) | (sec) |
| | 1800.0 | 1799.5 | 1800.5 | 1800.0 | 1800.0 | 0.07 |
| | 1800.0 | 1799.5 | 1800.5 | 1799.9 | 1799.9 | 0.06 |